

TOTAL CHLORINE

RANGE 0.0 TO 10.0 mg/L TEST STRIPS

480010

481110

BOTTLE OF 50 TEST STRIPS

30 TEST STRIPS

This patented (U.S. Patent No. 5620658) test is the sensitive and safe alternative to wet chemical chlorine testing.

Inexpensive
Reliable
Accurate results in less than three minutes.
Pocket sized
On site testing.
Ideal for water industry

No mechanical reader is required.
Reproducible results
Portable
Safe and easy to use.
No material safety data sheets required.

Available as bottle of 50 strips (480010) or 30 individual strips (481110).

| PART NUMBER | 480010 | 481110 |
|-------------|---|---|
| STEPS | 13 | 10 |
| PPM/mg/l | 0.0 TO 10.0 | 0.0 TO 10.0 |
| INCREMENTS | 0; 0.05; 0.1; 0.15; 0.2; 0.5; 0.8; 1.0; 4.0; 10.0 | 0; 0.05; 0.1; 0.15; 0.2; 0.5; 0.8; 1.0; 4.0; 10.0 |
| TEST TIME | 30 seconds | 30 seconds |

SenSafe™ TOTAL CHLORINE
 Free Chlorine + Monochloramines
CHECK™
 Patent Pending

TEST PROCEDURE:

METHOD A:
 Dip one test strip into a 250mL (8oz) water sample with constant, gentle back-and-forth motion for 5 seconds. Remove the strip and wait 30 seconds. Then, view through the aperture to match with closest METHOD A COLOR. Complete color match within 15 seconds.

TOTAL CHLORINE - METHOD A COLORS

| | | | | | | | |
|------------|-----|-----|-----|-----|-----|-----|------|
| PPM (mg/L) | 0.0 | 0.1 | 0.2 | 0.5 | 0.8 | 4.0 | 10.0 |
| | | | | | | | |

METHOD B:
 Dip one test strip into a 250mL (8oz) water sample with constant, gentle back-and-forth motion for 10 seconds. Remove the strip and wait 30 seconds. Then, view through the aperture to match with closest METHOD B COLOR. Complete color match within 15 seconds.

TOTAL CHLORINE - METHOD B COLORS

| | | | | | | | |
|------------|-----|------|-----|------|-----|-----|-----|
| PPM (mg/L) | 0.0 | 0.05 | 0.1 | 0.15 | 0.2 | 0.5 | 1.0 |
| | | | | | | | |

Contains 30 Test Strips Part Number 481110
 FOR BEST RESULTS, PLEASE FOLLOW INSTRUCTIONS CAREFULLY.

TEST PROCEDURES:

Method A Colors:

Method B Colors:

FOR OPTIMUM RESULTS use a 250 ml (1 cup) test sample. If the test sample is smaller, the results will read high (i.e., a 25 ml sample may read about 2 color blocks higher). A larger test sample (i.e., 500 ml or 1000 ml) will have a negligible impact on the results.

IMPORTANT: Store in dry, cool (below 30°C) place and away from direct sunlight. Use strips by printed "Use By" date.

National Primary Sanitation Drinking Water Regulations set forth by EPA recommend a Total Chlorine level of 4.0 PPM or less.

For technical assistance, call: 1-803-329-0162
REORDER NUMBER 481110
 MADE IN USA R119-TCL ©1999

Thermo Electron Station Road, Auchtermuchty, Fife, UK KY14 7DP
Tel: +44 (0)1337 828871 **Fax:** +44 (0)1337 828972
E-mail: sales.water.uk@thermo.com **Website:** <http://www.thermo.com>